


*DCV - 0 to 40 KV, ACV - 0 to 28 KV rms*

# **HIGH VOLTAGE PROBE**

**Model : HV-40**

*ISO-9001, CE, IEC1010*

---



---

**LUTRON ELECTRONIC**

***The Art of Measurement***

# HIGH VOLTAGE PROBE

Model : HV-40

SPECIFICATIONS		
Attenuate Ratio	1 : 1000.	
Input Impedance	Approx. 1000 M ohm.	
Output Impedance	Around 1.1 M ohm. The input impedance of external voltmeter should be 10 mega ohm.	
Max. Working Voltage	DCV	DC 40 KV
	ACV	Peak AC 40 KV or 28 KV rms ( depend which values is larger ).
Accuracy	DCV	1 KV to 20 KV - $\pm 1\%$ . 20 KV to 40 KV - $\pm 1.5\%$ .
	ACV	1 to 28 KV rms, 50/60 Hz - $\pm 5\%$ .
Temp Coefficient	Less than 200 ppm/°C.	
Operating Temperature	0 to 50 °C ( 32 to 122 °F ).	
Operating Humidity	Less than 80% RH.	
Cable Length	1 meter.	

## OPERATION

Connect the plugs to the volts ( Hi ) & com ( Lo ) input terminals of your voltmeter ( or Multimeter ). Select the desired range of voltmeter ( Attention : Do not use auto ranging ). Whenever possible, turn the high voltage source off before making any connections. Connect the HV probe common lead ( alligator clip ) to a good earth ground or reliable chassis ground.

## SAFETY PRECAUTION & WARNING !!!

- \* This high voltage probe must be used by the person who are trained only. Do not work alone when working with high voltage circuits & environment.
- \* For your own safety, inspect the probes for cracks & frayed or broken leads before each use. If any defects are noted, do not use the probes.
- \* Hands, shoes, floor & work bench must be dry. Avoid making measurements under humid, damp or other environmental conditions that might affect the safety of measurement situation.
- \* The ground connection must always be made before the probe tip comes into contact with the high voltage & must not be removed until after the probe tip has been removed from high voltage source.
- \* Do not attempt to take measurement from sources where the chassis or return lead is not ground.
- \* If possible, always turn the high voltage source off before connecting or disconnecting the probe.
- \* Before turning the high voltage on, make sure that no part of your body is in contact with the device under test.
- \* The probe body should be kept clean & free of any conductive contamination. Clean only the exterior probe body & cables. Use a soft cotton cloth lightly moistened with a mild solution of detergent & water. Do not allow any portion of the probe to be submerged at any time.